

Calibration Certificate

Calibration date: **05/05/2023**

Region: **DURHAM-POLICE**

Client

Name: Durham Regional Police Service
 Address: P.O. Box 911 480 Taunton Road
 City: Whitby, Ontario L1N 0B8

Scale

Manuf.: HAENNI

Model: WL-101

Serial Number: **35067**

Temperature

Start: **16 °C**

End: **16 °C**

X

Calibration required

No calibration required

RESULTS

Before calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
	▲	▼	▲	▼
0 kg	0	0	0	0
500 kg	509	522	9	22
1000 kg	998	996	-2	-4
1500 kg	1502	1507	2	7
2000 kg	1987	1992	-13	-8
2500 kg	2505	2492	5	-8
3000 kg	2994	2997	-6	-3
4000 kg	3980	3995	-20	-5
5000 kg	4975	4980	-25	-20
6000 kg	5980	5981	-20	-19
7000 kg	6984	6976	-16	-24
8000 kg	7990	7980	-10	-20
9000 kg	8974	8975	-26	-25
10,000 kg	9964	9964	-36	-36

▲ Increasing load

▼ Decreasing load

After calibration*

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
	▲	▼	▲	▼
0 kg	-	-	-	-
500 kg	-	-	-	-
1000 kg	-	-	-	-
1500 kg	-	-	-	-
2000 kg	-	-	-	-
2500 kg	-	-	-	-
3000 kg	-	-	-	-
4000 kg	-	-	-	-
5000 kg	-	-	-	-
6000 kg	-	-	-	-
7000 kg	-	-	-	-
8000 kg	-	-	-	-
9000 kg	-	-	-	-
10,000 kg	-	-	-	-

* Following a verification of the scale, no calibration will be done if the test results are within the manufacturer's tolerances.

Technician's signature:



Company:

Traffic Technology 2000

1-800-363-6224