

# Calibration Certificate

Calibration date: 05/05/2023

Region: DURHAM-POLICE

**Client**

Name: Durham Regional Police Service  
 Address: P.O. Box 911 480 Taunton Road  
 City: Whitby, Ontario L1N 0B8

**Scale**

Manuf.: HAENNI

Model: WL-101

Serial Number: 31682

**Temperature**

Start: 16.57

End: 16.57

X

Calibration required

No calibration required

## RESULTS

Before calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
	▲	▼	▲	▼
0 kg	0	0	0	0
500 kg	503	496	3	-4
1000 kg	988	995	-12	-5
1500 kg	1490	1486	-10	-14
2000 kg	2002	1998	2	-2
2500 kg	2504	2504	4	4
3000 kg	3002	2991	2	-9
4000 kg	4010	4014	10	14
5000 kg	5014	5005	14	5
6000 kg	6007	6009	7	9
7000 kg	7026	7011	26	11
8000 kg	8020	8014	20	14
9000 kg	9020	9027	20	27
10,000 kg	10011	10011	11	11

▲ Increasing load

▼ Decreasing load

After calibration\*

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
	▲	▼	▲	▼
0 kg	-	-	-	-
500 kg	-	-	-	-
1000 kg	-	-	-	-
1500 kg	-	-	-	-
2000 kg	-	-	-	-
2500 kg	-	-	-	-
3000 kg	-	-	-	-
4000 kg	-	-	-	-
5000 kg	-	-	-	-
6000 kg	-	-	-	-
7000 kg	-	-	-	-
8000 kg	-	-	-	-
9000 kg	-	-	-	-
10,000 kg	-	-	-	-

\* Following a verification of the scale, no calibration will be done if the test results are within the manufacturer's tolerances.

Technician's signature:



Company:

Traffic Technology 2000

1-800-363-6224