

Calibration Certificate

Calibration date: **08/14/2020**
 Region: **DURHAM-POLICE**

Client

Name: Durham Regional Police Service
 Address: P.O. Box 911 480 Taunton Road
 City: Whitby, Ontario L1N 0B8

Scale

Manuf.: HAENNI Model: WL-101 Serial Number: **32196**

Temperature

Start: **30 °C** End: **30 °C**

Calibration required
 No calibration required

RESULTS

Before calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
0 kg	0	-1	0	-1
500 kg	497	504	-3	4
1000 kg	996	999	-4	-1
1500 kg	1492	1484	-8	-16
2000 kg	1984	1975	-16	-25
2500 kg	2484	2471	-16	-29
3000 kg	2986	2976	-14	-24
4000 kg	3979	3953	-21	-47
5000 kg	4977	4945	-23	-55
6000 kg	5985	5958	-15	-42
7000 kg	6985	6961	-15	-39
8000 kg	7983	7961	-17	-39
9000 kg	8987	8975	-13	-25
10,000 kg	10006	10006	6	6

After calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
0 kg				
500 kg				
1000 kg				
1500 kg				
2000 kg				
2500 kg				
3000 kg				
4000 kg				
5000 kg				
6000 kg				
7000 kg				
8000 kg				
9000 kg				
10,000 kg				

Increasing load

Decreasing load

* Following a verification of the scale, no calibration will be done if the test results are within the manufacturer's tolerances.

Technician's Signature:



Company:

Traffic Technology 2000

1-800-363-6224