

# Calibration Certificate

Calibration date: **06/16/2021**  
 Region: **DURHAM-POLICE**

**Client**

Name: Durham Regional Police Service  
 Address: P.O. Box 911 480 Taunton Road  
 City: Whitby, Ontario L1N 0B8

**Scale**

Manuf.: HAENNI                      Model: WL-101                      Serial Number: **35066**

**Temperature**

Start: **15 °C**                      End: **15 °C**

Calibration required  
 No calibration required

### RESULTS

Before calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
0 kg	0	0	0	0
500 kg	504	486	4	-14
1000 kg	990	975	-10	-25
1500 kg	1483	1465	-17	-35
2000 kg	1996	1969	-4	-31
2500 kg	2498	2477	-2	-23
3000 kg	3013	2985	13	-15
4000 kg	4008	3985	8	-15
5000 kg	5005	4989	5	-11
6000 kg	6021	5990	21	-10
7000 kg	7014	7001	14	1
8000 kg	8006	7984	6	-16
9000 kg	9004	8995	4	-5
10,000 kg	9998	9998	-2	-2


After calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
0 kg				
500 kg				
1000 kg				
1500 kg				
2000 kg				
2500 kg				
3000 kg				
4000 kg				
5000 kg				
6000 kg				
7000 kg				
8000 kg				
9000 kg				
10,000 kg				

Increasing load

Decreasing load

\* Following a verification of the scale, no calibration will be done if the test results are within the manufacturer's tolerances.

Technician's Signature: 

Company:

*Traffic Technology 2000*

1-800-363-6224