



Infection Prevention and Control (IPAC) Lapse Report

Initial Report

Premises/facility under investigation:

Daniels Dental

Address:

Unit #1 - 1525 Dundas St E
Whitby, ON
L1N 2K6

Type of premises/facility: Dental Clinic

IPAC Lapse Details

Date board of health became aware of potential IPAC lapse: 2021-01-26

Date of initial report posting: 2021-02-01

How the potential IPAC lapse was identified: Complaint

Summary description of the IPAC lapse by complainant:

Concerned with improper sterilization.

IPAC Lapse Investigation

Did the IPAC lapse involve a member of a regulatory college? Yes

If yes, was the issue referred to the regulatory college? Yes

Please provide name of the regulatory college(s) and date(s) referred:

Royal College of Dental Surgeons of Ontario on 2021-01-29.

Were any corrective measures recommended and/or implemented? Yes

Please provide further details/steps:

- Ensure all staff responsible for reprocessing dental/medical devices and equipment are trained on cleaning, disinfection and sterilization.
- Ensure all sterilization parameters (physical, chemical and biological) are verified and recorded for each cycle load.
- Ensure expired items are discarded.

- Ensure each sterilizer is tested with a spore test/biological indicator for every cycle type setting used each day.

Date any order(s) or directive(s) were issued to the owner/operator:

2021-01-27 and 2021-01-29.

Final Report

Date of final report posting: 2021-08-24

Date all corrective measures were confirmed: 2021-03-18

Brief description of corrective measures taken: All items were in compliance as per the corrective measures indicated above. The extent of the above corrective measures was considered minimal.

Final Report Comments and Contact Information

Any additional comments: Operator education and resources were provided.

If you have any further questions, please contact the Durham Health Connection Line:

Telephone: 1-800-841-2729, or [email](#) us.