



Infection Prevention and Control (IPAC) Lapse Report

Initial Report

Premises/facility under investigation:

Daniels Dental

Address:

Unit #1 - 1525 Dundas St E
Whitby, ON
L1N 2K6

Type of premises/facility: Dental Clinic

IPAC Lapse Details

Date board of health became aware of potential IPAC lapse: 2021-01-26

Date of initial report posting: 2021-02-01

How the potential IPAC lapse was identified: Complaint

Summary description of the IPAC lapse by complainant:

Concerned with improper sterilization.

IPAC Lapse Investigation

Did the IPAC lapse involve a member of a regulatory college? Yes

If yes, was the issue referred to the regulatory college? Yes

Please provide name of the regulatory college(s) and date(s) referred:

Royal College of Dental Surgeons of Ontario on 2021-01-29.

Were any corrective measures recommended and/or implemented? Yes

Please provide further details/steps:

- Ensure all staff responsible for reprocessing dental/medical devices and equipment are trained on cleaning, disinfection and sterilization.
- Ensure all sterilization parameters (physical, chemical and biological) are verified and recorded for each cycle load.
- Ensure expired items are discarded.

- Ensure each sterilizer is tested with a spore test/biological indicator for every cycle type setting used each day.

Date any order(s) or directive(s) were issued to the owner/operator:

2021-01-27 and 2021-01-29.

Initial Report Comments and Contact Information

Any additional comments: Issued inspection reports detailing directives. Operator education and resources were provided.

If you have any further questions, please contact the Durham Health Connection Line:

Telephone: 1-800-841-2729, or [email](#) us.