

# Calibration Certificate

Calibration date: **06/14/2019**  
 Region: **DURHAM-POLICE**

**Client**

Name: Durham Regional Police Service  
 Address: P.O. Box 911 480 Taunton Road  
 City: Whitby, Ontario L1N 0B8

**Scale**

Manuf.: HAENNI                      Model: WL-101                      Serial Number: **35067**

**Temperature**

Start: **20 °C**                      End: **20 °C**

Calibration required  
 No calibration required

### RESULTS

Before calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
0 kg	0	0	0	0
500 kg	509	499	9	-1
1000 kg	1003	991	3	-9
1500 kg	1501	1491	1	-9
2000 kg	2003	1987	3	-13
2500 kg	2504	2484	4	-16
3000 kg	3004	2981	4	-19
4000 kg	4003	3987	3	-13
5000 kg	4993	4977	-7	-23
6000 kg	5987	5981	-13	-19
7000 kg	6990	6981	-10	-19
8000 kg	7977	7976	-23	-24
9000 kg	8972	8973	-28	-27
10,000 kg	9961	9961	-39	-39

After calibration

Load indicated on the scale	Load indicated on the test load cell (kg)		Difference (kg)	
0 kg				
500 kg				
1000 kg				
1500 kg				
2000 kg				
2500 kg				
3000 kg				
4000 kg				
5000 kg				
6000 kg				
7000 kg				
8000 kg				
9000 kg				
10,000 kg				

Increasing load

Decreasing load

\* Following a verification of the scale, no calibration will be done if the test results are within the manufacturer's tolerances.

Technician's Signature:



Company:

*Traffic Technology 2000*

1-800-363-6224